AND STATE OF THE PARTY OF THE P

Republic of the Philippines WESTERN MINDANAO STATE UNIVERSITY

Zamboanga City Telefax.: 062-991-7875

Email: <u>bacsecretariate@wmsu.edu.ph</u>

Website: www.wmsu.edu.ph

BID BULLETIN

PhilGEPS Reference No.: 7565260

Solicitation: PR-21-02-062

Project Name: Supply and Delivery of Various

Laboratory Equipment for the College of

Engineering of the University

(ABC: **Php. 4,900,100.00**)

After the Pre-Bid Conference held April 6, 2021, Tuesday 10:00 AM at the BAC Office, Western Mindanao State University, Normal Road, Zamboanga City, the following changes/clarifications has been agreed upon the bidders, the enduser, the TWG and BAC members;

A. CHANGES/CLARIFICATIONS:

То
Compact Muffle Furnace Specifications: * Chamber Size: 200 x 200 x 200 mm (8" x 8" x 8") * Working Temperature: 1300C (continuous) * Maximum Temperature: 1400C * Temperature Control: 30 steps programmable and PID automatic control * Heating Rate: 0-15C/min (suggestion: - 10C) * Temperature Accuracy: +/- 1C * Heating Element: Silicon Carbide or Ceramic * Working Voltage: AC 220V, single Phase, 60/50 Hz * With Accessories * With on-site Training</th

Item #2

Tensile Strength Testing Machine (500kg capacity)

Specifications:

- * Capacity: Maximum 5000N/500kg/1100 pounds
- * Grade: Semi-Automatic
- * Power: 220V,60hz, Single or 3 phase
- * With Accessories
- * With on-site Training

Tensile Strength Testing Machine

Specifications:

- * Capacity: Maximum 5000N or it's equivalent
- * Grade: Semi-Automatic
- * Power: 220V,60hz, Single or 3 phase
- * With Accessories
- * With on-site Training

Item #4

Laser Cutter for 3D Structure (1000W)

Specification:

- * Working Area: 1500mm* 3000mm (Saw tooth)
- * Cutting Speed: 100-300mm/sec
- * Laser Power: 300W/500W
- * Position Accuracy: ±0.05mm/m
- * Position Speed: 20m/min
- * Min. Line Width: ±0.02mm
- * Cutting thickness: ?4mm(stainless steel) ?6mm(carbon steel)
- * Max. Idle Running Speed: 1400mm/s
- * Transmit system: Dual Rack & Pinion type or its equivalent
- * Computer Interface: USB
- * Voltage: 220V/60HZ,3PH
- * With Instillation and commissioning
- * With Accessories and Hard Copy of Operating Manual in English

Laser Cutter for 3D Structure Specification:

* Working Area: 1200mm* 2500mm

* Cutting Speed: 100-300mm/sec

* Laser Power: 300W

* Position Accuracy: ±0.05mm/m

- * Position Speed: 20m/min * Min. Line Width: ±0.02mm
- * Cutting thickness: 1mm to 3mm
- * Max. Idle Running Speed: 1400mm/s
- * Transmit system: Dual Rack & Pinion type or its equivalent
- * Computer Interface: USB
- * Voltage: 220V/60HZ,3PH
- * With Installation and commissioning
- * With Accessories and Hard Copy of Operating Manual in English

B. DELIVERY PERIOD:

From	То
30 calendar days	90 calendar days from the receipt of PO/NTP

Reminder to Bidders:

- 1. All bidders are required to post a Bid Security in a form of Cash, Manager's or Cashier's Check or Bid Securing Declaration. Personal Check is NOT allowed.
- 2. Bidder must submit brochure/catalogue indicating the brand name and model of bid items.

- 3. Bidder must be an authorized reseller/distributor of the offered items and authorized by the manufacturer to provide/give technical support. Must attach resellership/distributorship certification.
- 4. Bidder's technical support trainors are authorized by the manufacturer to provide technical training and support of the products offered. Must attach supporting documents/certification to support such claims.
- 5. Single Largest Completed Contract (SLCC) shall be at least 50% of the cumulative Approved Budget for the Contract (ABC) of the item/s that correspond to the bid proposals.
- 6. Bid Submission will be on or before May 6, 2021 @ 10:00 A.M. through **Manual Submission.** We encourage bidders to observe proper lead time in the transmission of bids to guarantee that bids proposals are received by the WMSU BAC before the scheduled time of opening.
- 7. The Opening of the bid will be on <u>May 6, 2021 @ 10:00 A.M.</u> at the BAC Office, Western Mindanao State University, Ground Floor Executive Building, Normal Road, Baliwasan, Zamboanga City.

For further inquiries, you may coordinate with the Bids and Awards Committee Secretariat at Tel. No. 991-7875/0905-6310738.

Please be guided accordingly.

SGD

MA. LOURDES B. ALBA

BAC Chair