



Republic of the Philippines
WESTERN MINDANAO STATE UNIVERSITY
Zamboanga City
Telephone No.: 062-991-7875
www.wmsu.edu.ph

INVITATION TO BID

The Western Mindanao State University, through its Bids and Awards Committee (BAC), is inviting PhilGEPS registered suppliers to apply for eligibility and to submit bids for the item mentioned hereunder:

- 1.) 1 Unit **Medical Emergency Cart**
 - * Made of painted metal sheet
 - * With standard drawers and cabinet doors
 - * IV hook with rubber bumper
 - * 4 caster wheels with lock

- 2.) 5 Units **Modern Hospital Bed with Mattress**

specification:

Bed

 - * Overall dimension: at least L2100 x W970 x H480 mm
 - * Back rest tilting: 80 degree +/-5 degrees
 - * Knee board tilting: 40 degree +/- 5 degrees

Mattress

 - * Medical grade mattress

- 3.) 1 Unit **Light Weight, Bi-Phasic Defibrillator with Manual**

Specification:

 - * Low energy biphasic defibrillator monitor with recorder
 - * Shocks Driver: 2 joules to 200 joules
 - * Monitor ECG through paddles, pad and monitoring electrodes
 - * Defibrillate through pads and paddles
 - * Range: 25 to 150 ohms
 - * Built-in 50 mm strip Printer
 - * High resolution of more than 8 inches
 - * Viewing monitoring parameters: ECG, SpO2, NIBP, et CO2
 - * with adult and pediatric paddles
 - * Printing: at least 55 events
 - * Power: 220 V, 60 Hz
 - * Can be operated with battery
 - * With professional AED Trainer
 - * with on-site training
 - * With carrying case

- 4.) 1 Unit **Baby Intensive Care Incubator**

Technical specifications:

 - * Power Requirement: </- 400 VA
 - * Temp control mode: Air swervo control

- * Air temp control range: 20C - 37C
- * Temp override mode: 20C - 39C
- * Temp fluctuation: </-0.5C
- * Mattress temp: uniformity: </-0.8C
- * Temp rise time: </-45 min
- * Noise level within hood environment: </-55dB

Spefication of Hood:

- * Mattress to hoo: 450 mm
- * Observation to hood: Two incline
- * Operate port: 6 (2 ins ports)
- * Tubing grommets: 4
- * Front access panel: at least 826 x 310 mm
- * Mattress tray size: 630 x 360 mm

Alarms:

- * Power failure alarm
- * Fan Failure alarm
- * Sensor failure alarm
- * High airv temperature alarm: +3.0 c
- * Low air temperature alarm: -3.0 c
- * Over air temperature alarm: </- 38.0C; </-40.0C

5.) 1 Unit

Trauma Bag

Technical Specifications:

- * Height: 14"
- * width: 12"
- * Material: condura/Structural foam
- * Standard Color
- * with oxygen port and inclusions

6.) 4 Units

Collapsible Ambulance Stretcher

Specifications:

- * With two heavy duty seat belts to secure patients
- * Frame is made with high quality aluminum alloy materials
- * Can be converted to seating position
- * Reclining backrest
- * Folding legs controlled with safety handles
- * Equipped with safety lock to ensure stretcher is stable while in use
- * With upholstered mattress
- * Mounted on 4 casters with brakes and 2 additional rolling casters
- * Back angle: 0-55 degree (+/-5 degrees)
- * Maximum weight capacity: 200 kg

7.) 1 Unit

Sterilizer for the use of Dental Clinic

Specifications:

- * External dimensions: (W x D x H): at least 445mm x 532 mm x 428 mm
- * Chamber dimensions: at least 240mm x 384 mm
- * Heavy duty
- * Power Requirement: 220 V. 60Hz, Single phase

The criteria to be used for the eligibility check of the prospective bidders, examination and evaluation of bids, post-qualification and all matters relevant to this procurement shall be in accordance with Republic Act. No. 9184 (The Government Procurement Reform Act) and its Implementing Rules and Regulations.

Interested bidders may obtain further information from WMSU BAC Secretariat regarding the checklist of eligibility and technical requirements.

IMPORTANT NOTICE FOR BIDDERS:

1. Bidding papers shall be available upon payment of a non-refundable fee of One Thousand Pesos Only (Php.1,000.00). Eligible bidders that will download the bidding paper from the PhilGEPS website shall pay the said fee before the submission of their bids. (Please attached the Xerox copy of the Official Receipt)
2. All bidders' are required to post a Bid Security, at least Two (2%) Percent of the ABC in the form of Cash, Cashier's Check or Manager's Check or may submit Bid Securing Declaration. Bids without Bid Security will not be considered.
3. The Bidder shall prepare an original of the Eligibility Documents, Technical Proposal, Financial Proposal and clearly mark each "ORIGINAL – ELIGIBILITY DOCUMENTS", "ORIGINAL - TECHNICAL PROPOSAL", and "ORIGINAL – FINANCIAL PROPOSAL", respectively. Bidders shall submit three (3) copies of each bidding documents- "Original", Copy 1", and Copy 2". State the unit price of each item and the total bid price and also state the shortest time of delivery and submit your quotation duly signed by your representative in a sealed envelope.
4. Bid Submission will be on or before **April 8, 2021 at 10:00 A.M.** through **Manual Submission**.
5. Bid opening shall be on **April 8, 2021** at 10:00 A.M. at BAC Office, Ground Floor Executive Building, WMSU, Zamboanga City, Philippines, 7000. Bids will be opened in the presence of the Bidders representatives who choose to attend at the address above. Late bids shall not be accepted.
6. Price validity shall be for a period of 120 calendar days.
7. Bidders shall submit original brochures showing certifications of the product being offered.
8. Warranty shall be for a period of Six (6) months for supplies and materials. One (1) year for equipment, from date of acceptance by WESTERN MINDANAO STATE UNIVERSITY.
9. Bids received in excess of the ABC shall be automatically rejected at Bid Opening.

The WESTERN MINDANAO STATE UNIVERSITY reserves the right to reject any or all Bids and to accept the bid most advantageous to the government, and to award the contract by lot, if warranted.