



INVITATION TO BID

The Western Mindanao State University, through its Bids and Awards Committee (BAC), is inviting PhilGEPS registered suppliers to apply for eligibility and to submit bids for the item mentioned hereunder:

Name of the Project: **PR 24-05-262 Rebid Procurement of Dental Chair and Autoclave for the Dental Clinic of University Health Services**

Approved Budget for the Contract (ABC): **Php. 390,000.00**

Bid Document: **Php 500.00**

Delivery Period: **Thirty (30) Calendar Days**

1. *One (1) unit of DENTAL CHAIR*

Specifications:

overall size: at least 1900mm x 920mm x 200mm

base: at least 880mm x 600mm

Maximum Height of Seat Cushion from Ground: greater than or equal to 450mm

Maximum Angle of Seat Cushion Back: less than or equal to 10 degrees

Range of Backward Angle: 105 degrees - 170 degrees

Expansion Range of Headrest: 0-120mm

Load Capacity of the E60 Dental Unit: less than or equal to 1350N

Classification of Medical Instruments: Class II

Noise: less than or equal to 42dB

Air Compressor Specification (silent type)

Power Voltage- 220V AC TOH2

Power - 850W

Gas Production Volume 100L/min (0.4mpa)

Pressure Summit - 0.8 mpa

Rotating speed - 1400/min

Cubage - 32L

Standard Configuration

-Dental Chair

-Dental Stool

-Dentist Console

-Assistant Console

-3 way Syringe - 2 sets

-Weak Suction Saliva ejector Valve - 1 set

-Strong Suction Saliva Ejector valve - 1 set

-Intra Oral Camera - 1 set

-Dental X-ray Viewer

-Foot Switch - 1 set

-Quantitative Water supply control system

Cuspidor Bowl Rinse and mouth rinse - 1 set

8=holes Led sensor light

Accessories:

4 holes high speed hand piece

4 holes low speed hand piece

Ultrasonic scaler

Light cure machine

2. *One (1) unit of Autoclave (18 Liters)*

Specification:

Machine size: at least (L)56 x (W)31.5 x (H)36.5C

Sterilize temperature 121/134 Degree celsius

Sterilize Pressure (Mpa) 0.12Mpa holding



Time 121 degree / 30 min 134 degree /10 min
No dry Funtion Power: 1100VAC
Chamber size: Stainless steel 18 liters
Safety: Heating-over protection: Safety Valve
TEMP CONTROL DESIGN: Chip Control
Electric Current 220-240VAC, 60 Hz

Accessories

Tray Basket, Tray Holder
Sterilization Tray, steam exhaust pipe

Delivery Period: 30 Days

The criteria to be used for the eligibility check of the prospective bidders, examination and evaluation of bids, post-qualification and all matters relevant to this procurement shall be in accordance with Republic Act. No. 9184 (The Government Procurement Reform Act) and its Implementing Rules and Regulations.

Interested bidders may obtain further information from WMSU BAC Secretariat regarding the checklist of eligibility and technical requirements.

IMPORTANT NOTICE FOR BIDDERS:

1. Bidding papers may be acquired starting September 13 to September 24, 2024 from the BAC Office or download from website (www.wmsu.edu.ph). The WESTERN MINDANAO STATE UNIVERSITY shall allow the bidder to present its proof of payment for the Bidding Documents fees before the submission of their bids, pursuant to the latest Guidelines issued by the GPPB, in the amount of Five Hundred Pesos (Php.500.00). (Please attached the machine copy of the Official Receipt)
2. All bidders are required to post a Bid Security, at least Two (2%) Percent of the ABC in the form of Cash, Cashier's Check or Manager's Check or may submit Bid Securing Declaration. Bids without Bid Security will not be considered.
3. The Bidder shall prepare an original of the Eligibility Documents & Technical Components and original of Financial Proposal and clearly mark each "ORIGINAL – ELIGIBILITY DOCUMENTS and TECHNICAL COMPONENTS", and "ORIGINAL – FINANCIAL PROPOSAL", respectively. Bidders shall submit one (1) set of the first and second components of its bid. State the unit price of each item and the total bid price and also state the shortest time of delivery and submit your quotation duly signed by your representative in a sealed envelope.
4. Bid Submission will be on or before September 24, 2024 at 9:30 A.M. through Manual Submission.
5. Bid opening shall be on September 24, 2024 at 9:30 A.M. at BAC Office, 2nd Floor Administration Building, WMSU, Zamboanga City, Philippines, 7000. Bids will be opened in the presence of the Bidders representatives who choose to attend at the address above. Late bids shall not be accepted.
6. Price validity shall be for a period of 120 calendar days.
7. Bidders shall submit original brochures showing certifications of the product being offered.
8. Warranty shall be for a period of Six (6) months for supplies and materials. One (1) year for equipment, from date of acceptance by WESTERN MINDANAO STATE UNIVERSITY.
9. Bids received in excess of the ABC shall be automatically rejected at Bid Opening.

The WESTERN MINDANAO STATE UNIVERSITY reserves the right to reject any or all



Republic of the Philippines
WESTERN MINDANAO STATE UNIVERSITY

Normal Road, Baliwasan Zamboanga City

www.wmsu.edu.ph

Telephone: 062-991-1771 LOC 1003



Bids and to accept the bid most advantageous to the government, and to award the contract by lot, if warranted.