



Republic of the Philippines
WESTERN MINDANAO STATE UNIVERSITY
Zamboanga City
Telephone No.: 062-991-7875
www.wmsu.edu.ph

REQUEST FOR QUOTATION

The Western Mindanao State University, through its Bids and Awards Committee (BAC), is inviting PhilGEPS registered suppliers to apply for eligibility and to submit bids for the item mentioned hereunder:

Name of Project: **Supply and Delivery of Various Laboratory Equipment**

Approved Budget Cost: **Php 208,000.00**

Purchase Request No.: **PR 23-01-064**

Closing Date: **April 13, 2023 @ 10AM**

Description:

1.) One (1) unit of Table Top pH Meter

pH Range -2.00 to 16.00 pH

Resolution 0.01 pH

Accuracy ± 0.01 pH

Cal. Points Up to 5

Buffer Sets USA (1.68, 4.01, 7.00, 10.01, 12.45),

NIST (1.68, 4.01, 6.86, 9.18, 12.45)

ORP

Range ± 2000 mV

Rel. mV Range ± 2000 mV

Resolution 0.1 mV (± 199.9 mV) / 1 mV (beyond)

Accuracy ± 0.2 mV (± 199.9 mV) / ± 2 mV (beyond)

Temperature

Range (Meter) 0.0 to 100.0 oC / 32.0 to 212.0 oF

Resolution 0.1 oC / 0.1 oF

Accuracy ± 0.3 oC / ± 0.5 oF (0 to 70 oC)

Compensation ATC / MTC (0 to 100 oC) (pH only)

Calibration Offset in 0.1 oincrements;

offset range: ± 5.0 oC / ± 9.0 oF

Power Requirements 9 V DC adapter, 1.3 A (100/240 VAC, SMPS).

2.) One (1) unit of Non-Electric Autoclave

Material: Aluminum

Liquid Capacity: 41 Quarts

Pressure gauge Weights: 5, 10, 15 psi

32 Pint Jars

19 Quart Jars

Inside Diamter: 15.5 Inches

Inside Height: 13.5 inches

Size: 49 (H) x 57 (W) x 46 (D) cm.

3.) One (1) unit of Centrifuge Machine

SPECIFICATIONS:

Speed range(rpm): 0-4000rpm adjustable

Max Capacity(ml): 6*50ml (Rotor included : 12x15ml)

RCF range (xg): 0-2700xg adjustable

Noise(dBA): ≤ 58 dB(A)

Time range: 1-99minutes adjustable

Speed accuracy : ± 20 r/min

Dimension(mm): 460*350*265mm

Net weight(without rotor): 23KGS
Power supply: AC110~240V,50/60HZ,1A,0.04KW
Protection: Lid Lock, Over Speed, Imbalance Protection.

The criteria to be used for the eligibility check of the prospective bidders, examination and evaluation of bids, post-qualification and all matters relevant to this procurement shall be in accordance with Republic Act. No. 9184 (The Government Procurement Reform Act) and its Implementing Rules and Regulations.

Interested bidders are required to submit their valid and current Mayor's Permit and PhilGEPS Registration, upon the submission of quotation.

Award of contract shall be made to the lowest calculated and responsive bid, which complies with the necessary description as stated above and other terms and conditions stated in the price quotation form.

Any interlineations, erasures or overwriting shall be valid only if they are signed or initiated by the bidder or his/her duly authorized representative/s.

Submission of Quotation and eligibility documents is on or before April 13, 2023 at the BAC Office, Ground Floor Executive Building, Western Mindanao State University, Normal Road, Baliwasan, Zamboanga City. Open submission may be submitted manually, email (bacsecretariate@wmsu.edu.ph) or through facsimile at (062)991-7875. For inquiries, you may coordinate with the BAC Secretariat at telephone no. (062)991-7875.

REQUEST FOR QUOTATION

Western Mindanao State University

Quotation No.: _____

PR No.: PR-23-01-064

Please quote your lowest price on the item/s listed below, subject to the General Conditions on the page, stating the shortest time of delivery and submit your quotation duly signed by your representative not later than APR 18 2020 at 10:00 AM in the return envelope attached herewith. Any quotation submitted beyond this date will not be considered.


FREDELINO M. SAN JUAN, Ph.D.
BAC Chair

NOTE:

- 1 ALL ENTRIES MUST BE TYPEWRITTEN
- 2 DELIVERY PERIOD WITHIN 10 CALENDAR DAYS FROM THE RECEIPT OF THE PURCHASE ORDER.
- 3 WARRANTY SHALL BE FOR A PERIOD OF SIX (6) MONTHS FOR SUPPLIES AND MATERIALS. ONE (1) YEAR FOR EQUIPMENT, FROM DATE OF ACCEPTANCE BY WESTERN MINDANAO STATE UNIVERSITY
- 4 PRICE VALIDITY SHALL BE FOR A PERIOD OF 120 CALENDAR DAYS UPON RECEIPT OF THE PURCHASE ORDER
- 5 G-EPS REGISTRATION CERTIFICATE SHALL BE ATTACHED UPON SUBMISSION OF THE QUOTATION
- 6 BIDDERS SHALL SUBMIT ORIGINAL BROCHURES SHOWING CERTIFICATIONS OF THE PRODUCT BEING OFFERED

Item No.	Qty	Unit	Item and Description	Approved Budget for Contract (ABC)	Unit Cost	Total Cost
1.	1	Unit	Table Top pH Meter pH Range – 2.00 to 16.00 pH Resolution 0.01 pH Accuracy ±0.01 pH Cal. Points Up to 5 Buffer Sets USA (1.68, 4.01, 7.00, 10.01, 12.45) NIST (1.68, 4.01, 6.86, 9.18, 12.45) ORP Range ±2000mV Rel. mV Range ±2000mV Resolution 0.1mV (±199.9 mV)/ 1 mV (beyond) Accuracy ±0.2 mV (±199.9 mV)/ ±2 mV (beyond) Temperature Range (Meter) 0.0 to 100.0 oC/ 32.0 212.0 oF Resolution 0.1 oC/ 0.1 oF Accuracy ±0.3 oC/ ±0.5 oF (0 to 70 oC) Compensation ATC/ MTC (0 to 100 oC) (pH only) Calibration Offset in 0.1 increments; Offset range: ±5.0 oC/ ±9.0 oF Power Requirements 9 V DC adapter, 1.3 A (100/240 VAC, SMPS)	P 63,000.00		
2.	1	Unit	Non-electric Autoclave Material: Aluminum Liquid Capacity: 41 Quarts Pressure gauge Weights: 5,10,15 psi 32 Pint jars 19 Quart Jars Inside Diameter: 15.5 inches Inside Height: 13.5 inches Size: 49 (H) x 57 (W) x 46 (D) cm	P 75,000.00		
			<i>Note: For the Shared Service Facility (SSF) Office of the University.</i>			

PAGE 1 of 2

EPS Reference Number : _____
 EPS Solicitation Number : _____
 EPS Closing Date : _____

Total: _____
 Brand & Model : _____
 Delivery Period : _____
 Warranty : _____
 Price Validity : _____

After having carefully read and accepted your General Conditions, the foregoing are our price quotation for the items above indicated.

PhilGEPS Registration No.: _____
 Certificate Reference No.: _____

REYNANTE ESPIRITUSANTO / JORGE CONCEPCION / RALPH JUDE LLACUÑA
 Canvasser

 Printed Name/Signature

 Tel .No./Cellphone #

 Date

REQUEST FOR QUOTATION

Western Mindanao State University

Quotation No.: _____

PR No.: PR-23-01-064

Please quote your lowest price on the item/s listed below, subject to the General Conditions on the page, stating the

shortest time of delivery and submit your quotation duly signed by your representative not later than APR 18 2023 at 10:00 AM in the return envelope attached herewith. Any quotation submitted beyond this date will not be considered.

FREDELINO M. SAN JUAN, Ph.D.

BAC Chair

NOTE:

- 1 ALL ENTRIES MUST BE TYPEWRITTEN
- 2 DELIVERY PERIOD WITHIN 10 CALENDAR DAYS FROM THE RECEIPT OF THE PURCHASE ORDER.
- 3 WARRANTY SHALL BE FOR A PERIOD OF SIX (6) MONTHS FOR SUPPLIES AND MATERIALS. ONE (1) YEAR FOR EQUIPMENT, FROM DATE OF ACCEPTANCE BY WESTERN MINDANAO STATE UNIVERSITY
- 4 PRICE VALIDITY SHALL BE FOR A PERIOD OF 120 CALENDAR DAYS UPON RECEIPT OF THE PURCHASE ORDER
- 5 G-EPS REGISTRATION CERTIFICATE SHALL BE ATTACHED UPON SUBMISSION OF THE QUOTATION
- 6 BIDDERS SHALL SUBMIT ORIGINAL BROCHURES SHOWING CERTIFICATIONS OF THE PRODUCT BEING OFFERED

Item No.	Qty	Unit	Item and Description	Approved Budget for Contract (ABC)	Unit Cost	Total Cost
3.	1	Unit	Centrifuge Machine Specifications: Speed range (rpm): 0-4000rpm adjustable Max Capacity(ml): 6*50ml (Rotor included: 12x15ml) RCF Range (xg): 0-2700xg adjustable Noise (dBA): <=58dB(A) Time Range: 1-99minutes adjustable Speed accuracy: ±20r/min Dimension(mm): 460*350*265mm Net Weight (without rotor): 23kgs Power Supply: AC110~240V, 50/60Hz, 1A, 0.04KW Protection: Lid Lock, Over Speed, Imbalance Protection	P 70,000.00		
			<i>Note: For the Shared Service Facility (SSF) Office of the University.</i>			

PAGE 2 of 2

EPS Reference Number : _____
 EPS Solicitation Number : _____
 EPS Closing Date : _____

Total: _____
 Brand & Model : _____
 Delivery Period : _____
 Warranty : _____
 Price Validity : _____

After having carefully read and accepted your General Conditions, the foregoing are our price quotation for the items above indicated.

PhilGEPS Registration No.: _____
 Certificate Reference No.: _____

REYNANTE ESPIRITUSANTO / JORGE CONCEPCION / RALPH JUDE LLACUÑA
 Canvasser

 Printed Name/Signature

 Tel .No./Cellphone #

 Date