



REQUEST FOR QUOTATION

The Western Mindanao State University, through its Bids and Awards Committee (BAC), is inviting PhilGEPS registered suppliers to apply for eligibility and to submit bids for the item mentioned hereunder:

Name of Project: **Procurement of Table Top PH Meter for the College of Agriculture of the University**

Approved Budget Cost: **Php 63,000.00**

Purchase Request No.: **PR 24-06-271**

Closing Date: **June 28, 2024 @ 9:30 AM**

Description:

1. One (1) unit of TABLE TOP PH METER

pH Range -2.00 to 16.00 pH

Resolution 0.01 pH

Accuracy ± 0.01 pH

Cal. Points Up to 5

Buffer Sets USA (1.68, 4.01, 7.00, 10.01, 12.45),

NIST (1.68, 4.01, 6.86, 9.18, 12.45)

ORP

Range ± 2000 mV

Rel. mV Range ± 2000 mV

Resolution 0.1 mV (± 199.9 mV) / 1 mV (beyond)

Accuracy ± 0.2 mV (± 199.9 mV) / ± 2 mV (beyond)

Temperature

Range (Meter) 0.0 to 100.0 oC / 32.0 to 212.0 oF

Resolution 0.1 oC / 0.1 oF

Accuracy ± 0.3 oC / ± 0.5 oF (0 to 70 oC)

Compensation ATC / MTC (0 to 100 oC) (pH only)

Calibration Offset in 0.1 oincrements;

offset range: ± 5.0 oC / ± 9.0 oF

Power Requirements 9 V DC adapter, 1.3 A (100/240 VAC, SMPS).

to be Deliver at College of Agriculture, San Ramon Campus, Zamboanga City

The criteria to be used for the eligibility check of the prospective bidders, examination and evaluation of bids, post-qualification and all matters relevant to this procurement shall be in accordance with Republic Act. No. 9184 (The Government Procurement Reform Act) and its Implementing Rules and Regulations.

Interested bidders are required to submit their valid and current Mayor's Permit and PhilGEPS Registration, upon the submission of quotation.

Award of contract shall be made to the lowest calculated and responsive bid, which complies with the necessary description as stated above and other terms and conditions stated in the price quotation form.

Any interlineations, erasures or overwriting shall be valid only if they are signed or initiated by the bidder or his/her duly authorized representative/s.

Submission of Quotation and eligibility documents is on or before **June 28, 2024** at the BAC Office, 2nd Floor, Admin Building, Western Mindanao State University, Normal Road, Baliwasan, Zamboanga City. Open submission may be submitted manually, email (bacsecretariate@wmsu.edu.ph) or through facsimile at (062)991-7875.

For inquiries, you may coordinate with the BAC Secretariat at telephone no. (062)991-7875

REQUEST FOR QUOTATION

Western Mindanao State University

Quotation No.: _____

PR No.: **24-06-271**

Please quote your lowest price on the item/s listed below, subject to the General Conditions on the page, stating the shortest time of delivery and submit your quotation duly signed by your representative not later than JUN 28 2024 at **9:30 A.M.** in the return envelope attached herewith. Any quotation submitted beyond this date will not be considered.

JOSELITO D. MADROÑAL, DPA
BAC Chairperson for GOODS

- NOTE:**
- 1 ALL ENTRIES MUST BE TYPEWRITTEN
 - 2 DELIVERY PERIOD _____ CALENDAR DAYS UPON RECEIPT OF THE PURCHASE ORDER.
 - 3 WARRANTY SHALL BE FOR A PERIOD OF SIX (6) MONTHS FOR SUPPLIES AND MATERIALS. ONE (1) YEAR FOR EQUIPMENT, FROM DATE OF ACCEPTANCE BY WESTERN MINDANAO STATE UNIVERSITY
 - 4 PRICE VALIDITY SHALL BE FOR A PERIOD OF 120 CALENDAR DAYS UPON RECEIPT OF THE PURCHASE ORDER
 - 5 G-EPS REGISTRATION CERTIFICATE SHALL BE ATTACHED UPON SUBMISSION OF THE QUOTATION
 - 6 BIDDERS SHALL SUBMIT ORIGINAL BROCHURES SHOWING CERTIFICATIONS OF THE PRODUCT BEING OFFERED

Item No.	Qty	Unit	Item and Description	Approved Budget for the Contract (ABC)	Unit Cost	Total Cost
1.	1	unit	<p>TABLE TOP PH METER <i>pH Range -2.00 to 16.00 pH</i> <i>Resolution 0.01 pH</i> <i>Accuracy ±0.01 pH</i> <i>Cal. Points Up to 5</i> <i>Buffer Sets USA (1.68, 4.01, 7.00, 10.01, 12.45),</i> <i>NIST (1.68, 4.01, 6.86, 9.18, 12.45)</i> ORP <i>Range ±2000 mV</i> <i>Rel. mV Range ±2000 mV</i> <i>Resolution 0.1 mV (±199.9 mV) / 1 mV (beyond)</i> <i>Accuracy ±0.2 mV (±199.9 mV) / ±2 mV (beyond)</i> Temperature <i>Range (Meter) 0.0 to 100.0 oC / 32.0 to 212.0 oF</i> <i>Resolution 0.1 oC / 0.1 oF</i> <i>Accuracy ±0.3 oC / ±0.5 oF (0 to 70 oC)</i> <i>Compensation ATC / MTC (0 to 100 oC) (pH only)</i> <i>Calibration Offset in 0.1 oincrements;</i> <i>offset range: ±5.0 oC / ±9.0 oF</i> <i>Power Requirements 9 V DC adapter, 1.3 A</i> <i>(100/240 VAC, SMPS).</i></p> <p><i>to be Deliver at College of Agriculture, San Ramon Campus, Zamboanga City</i></p>	₱ 63,000.00		
			<i>Note: for the College of Agriculture of the University</i>			

PAGE 1 Of 1

EPS Reference Number : _____	Total: _____
EPS Solicitation Number : _____	Brand & Model : _____
EPS Closing Date : _____	Delivery Period : _____
	Warranty : _____
	Price Validity : _____

After having carefully read and accepted your General Conditions, the foregoing are our price quotation for the items above indicated.

PhilGEPS Certificate No.: _____

Certificate Reference No.: _____

REY ESPIRITUSANTO / JORGE CONCEPCION / RALPH JUDE LLACUÑA
Canvasser

Printed Name/Signature

Tel .No./Cellphone #

Date